



No. 2011 - 051

Safety Certificate

- Manufacturer **KOREA FLOWMETER IND.CO.,LTD**
- Address **#40, Wonchang-ro 64 beon-gil, Seo-gu, Incheon, South Korea.**

The following product fabricated by the above manufacturer has been certified for the use of the Safety Certification Mark as the product was assessed and tested in accordance with the article 34 of the Occupational Safety & Health Act and the article 58 (4) 4 of the Enforcement Rule thereof.

- Product **Rotameter Indicator**
- Type(Model) **NMGE-S-EX**
- Rating **DC 24V**
- Type of Protection **Ex d IIC T5**
- Certification No. **11-GA2BO-051**
- Safety Certification Standards **the Article 34 of the Occupational Safety & Health Act**
- Certification condition **1. Only products produced at the above address are available
2. Each equipment shall be subjected to overpressure test of 1.5 times higher than reference pressure (1.2MPa : Enclosure, 1.02MPa : T/box)**
- Date of issue **May 24th, 2011**

Approved for issue on behalf of the president : **Gi-Hoi, Kim**
position : **General Manager**

Signature





**KOREA GAS SAFETY
CORPORATION**