

TEST REPORT



한국산업기술시험원
Korea Testing Laboratory

Report No. : 16-041445-01

Page of Pages : (1) / (5)



1. Client

Name : WISE CONTROL Inc.

Address : 2022, Deogyong-daero, Gilheung-gu, Yongin-si, Gyeonggi-do, 17097, Rep of Korea

Date of Receipt : 2016. 07. 14

2. Use of Report : To verify IP grade to IEC 60529

3. Test Sample

Description : Low Pressure Gauge

Manufacturer : WISE CONTROL Inc.

Model Name : P422

Serial Number : -

Remark : -



4. Date of Test : 2016. 07. 28. ~ 2016. 07. 29.

5. Test Standard/Method : IEC 60529: 2001

6. Testing Environment : Temperature : (25.1 ± 2.0) °C , Humidity : (63 ± 2) % R.H.

7. Test Results : Pass (IP65)

Note : 1. The test results contained apply only to the test sample(s) supplied by the client
2. This test report shall not be reproduced in full or in part without approval of the KTL in advance.

Affirmation	Tested by	 (Signature)	Technical Manager	 (Signature)
	Name : Chae Hui dong		Name : Min Yeong-seung	

The above test report is the accredited test result by Korea Laboratory Accreditation Scheme, which signed the ILAC-MRA.

2016. 08. 02

Korea Testing Laboratory

Accredited by KOLAS, Republic of KOREA




87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA Tel.+82-2-860-1537 Fax. +82-2-860-1549

FP202-03-03

※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.

<Contents>

1. Summary of Test	3
1.1 Test Standard	3
1.2 Test Sample	3
1.3 Test Environment	3
1.4 Remark	3
2. Results	4
2.1 Dust Test Conditions	4
2.2 Dust Test Contents	4
2.3 Water Test Conditions	4
2.4 Water Test Contents	4
3. List of Testing Equipments	4
4. Test Figures	5



1. Summary of Test

1.1 Test Standard

This test was conducted in accordance with "IEC 60529: 2001".

1.2 Test Sample

- Description : Low Pressure Gauge
- Model Name : P422
- Dimensions : $\Phi 101.3 \text{ mm} \times 50 \text{ mm}$
- Rating : -



[Fig. 1: Front]



[Fig. 2: Side]

1.3 Test Environment

- Temperature : $(25.1 \pm 2.0) ^\circ\text{C}$
- Humidity : $(63 \pm 2) \% \text{ R.H.}$
- Atmospheric Pressure : $(100.8 \pm 2.0) \text{ kPa}$

1.4 Remark

N/A

FP202-02-02

2. Results

Code Letters	IP	Conditions	Results
1st Characteristic numerals Against ingress of solid foreign objects	6	2.1 Dust Test Conditions <ul style="list-style-type: none"> Talcum powder(mesh) wire diameter: 50 μm Talcum powder(mesh) wire width: 75 μm Amount of talcum powder of the test chamber: 2 kg/m^3 2.2 Dust Test Contents <ul style="list-style-type: none"> Volume of the enclosures: About 250 cm^3 Reduction air pressure: -2.00 kPa (-200 mmH_2O) Flow rate: 0.00 LPM Extraction rate per hour: 0 volumes/h Test duration: 8 h 	Pass
2nd Characteristic numerals Against ingress of water with harmful effects	5	2.3 Water Test Conditions <ul style="list-style-type: none"> Internal diameter of the nozzle: 6.3 mm Delivery rate: 12.5 l/min \pm 5 % Core of the substantial stream: Circle of 40 mm diameter at 2.5 m distance from the nozzle Distance from nozzle to enclosure surface: 2.8 m 2.4 Water Test Contents <ul style="list-style-type: none"> Test duration: 3 min 	Pass

3. List of Testing Equipments

Equipment	Manufacture	Model	ICP No.	Date of Calibration	Calibration Laboratory
Thermo-hygrometer	TESTO	Testo 622	ICP20140892	2016. 05. 10	KTL
Vernia Caliper	MITUTOYO	CD-20APX	ICP20160207	2016. 05. 19	KTL
Flow Meter (Nozzle)	KOMETER	GA-101	ICP20150325	2016. 05. 13	DDHT
Flow Meter (Dust Chamber)	DWYER	RMA-13-SSV	ICP20150476	2016. 05. 11	KTL

FP202-02-02

4. Test Figures



[Fig. 3: IP 6X]



[Fig. 4: IP X5]

복사본 COPY

COPY 복사본

- End -

FP202-02-02



※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.